

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00780 Calib. Date: 03/09/2020 Calib. No.: 00030
Certification File No.: 00781 Cert. Date: 03/09/2020 Cert. No.: 00025
Linearity File No.: 00782 Lin. Date: 03/09/2020 Lin. No.: 00025
Solution File No.: 00788 Soln. Date: 06/11/2020 Soln. No.: 00195
Sequential File No.: 00788 File Date: 06/11/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064
Control Solution %: 0.100% Expires: 02/19/2022
Solution Control Lot: 20090 Bottle No.: 0019

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:14D		
Control 1 EC	0.099%	14:14D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:14D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15D		
Control 2 EC	0.098%	14:16D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:16D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:16D		
Control 3 EC	0.098%	14:17D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:17D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 06/11/2020