

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00780 Calib. Date: 03/09/2020 Calib. No.: 00030
Certification File No.: 00781 Cert. Date: 03/09/2020 Cert. No.: 00025
Linearity File No.: 00782 Lin. Date: 03/09/2020 Lin. No.: 00025
Solution File No.: 00790 Soln. Date: 07/07/2020 Soln. No.: 00196
Sequential File No.: 00790 File Date: 07/07/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064
Control Solution %: 0.100% Expires: 02/19/2022
Solution Control Lot: 20090 Bottle No.: 0144

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:06D		
Control 1 EC	0.097%	20:07D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	20:07D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:08D		
Control 2 EC	0.097%	20:08D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.097%	20:08D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 3 EC	0.097%	20:09D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:09D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 07/07/2020