

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE		
Calibration File No.:	00792	Calib. Date:	07/28/2020
Certification File No.:	00793	Cert. No.:	00031
Linearity File No.:	00794	Lin. Date:	07/28/2020
Solution File No.:	00797	Cert. No.:	00026
Sequential File No.:	00797	Lin. No.:	00026
		Soln. Date:	08/21/2020
		Soln. No.:	00198
		File Date:	08/21/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZH-0064
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0145

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:17D		
Control 1 EC	0.099%	16:18D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:18D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18D		
Control 2 EC	0.099%	16:19D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:19D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20D		
Control 3 EC	0.099%	16:20D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:20D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 08/21/2020