

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00792 Calib. Date: 07/28/2020 Calib. No.: 00031
Certification File No.: 00793 Cert. Date: 07/28/2020 Cert. No.: 00026
Linearity File No.: 00794 Lin. Date: 07/28/2020 Lin. No.: 00026
Solution File No.: 00801 Soln. Date: 08/31/2020 Soln. No.: 00199
Sequential File No.: 00801 File Date: 08/31/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064
Control Solution %: 0.100% Expires: 02/19/2022
Solution Control Lot: 20090 Bottle No.: 0202

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 07:48D | | |
| Control 1 EC | 0.096% | 07:48D | 33.8°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 07:48D | 33.8°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:49D | | |
| Control 2 EC | 0.095% | 07:50D | 33.8°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 07:50D | 33.8°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:50D | | |
| Control 3 EC | 0.096% | 07:51D | 33.8°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 07:51D | 33.8°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:51D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: LT. F. Pumphrey #3786

Badge No.: 3786

Date: 08/31/2020