

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00792	Calib. Date: 07/28/2020
Certification File No.:	00793	Calib. No.: 00031
Linearity File No.:	00794	Cert. Date: 07/28/2020
Solution File No.:	00808	Cert. No.: 00026
Sequential File No.:	00808	Lin. Date: 07/28/2020
		Lin. No.: 00026
		Soln. Date: 10/17/2020
		Soln. No.: 00201
		File Date: 10/17/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDZH-0064
Control Solution %:	0.100%			Expires:	02/19/2022
Solution Control Lot:	20090			Bottle No.:	0483

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:36D		
Control 1 EC	0.100%	09:37D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:37D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37D		
Control 2 EC	0.099%	09:38D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:38D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 3 EC	0.099%	09:39D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:39D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 10/17/2020