

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00792	Calib. Date: 07/28/2020
Certification File No.:	00793	Calib. No.: 00031
Linearity File No.:	00794	Cert. Date: 07/28/2020
Solution File No.:	00814	Cert. No.: 00026
Sequential File No.:	00814	Lin. Date: 07/28/2020
		Lin. No.: 00026
		Soln. Date: 12/25/2020
		Soln. No.: 00204
		File Date: 12/25/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0064
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0990

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:39S		
Control 1 EC	0.100%	14:40S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:40S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41S		
Control 2 EC	0.099%	14:41S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:41S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42S		
Control 3 EC	0.099%	14:43S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:43S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

W. F. Dwyer #3786

Badge No.: 3786

Date: 12/25/2020