

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00754	Calib. Date: 10/20/2019
Certification File No.:	00755	Calib. No.: 00029
Linearity File No.:	00756	Cert. Date: 10/20/2019
Solution File No.:	00768	Cert. No.: 00024
Sequential File No.:	00768	Lin. Date: 10/20/2019
		Lin. No.: 00024
		Soln. Date: 01/06/2020
		Soln. No.: 00187
		File Date: 01/06/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0063
Solution Control Lot:	19110	Expires: 04/09/2021
		Bottle No.: 0728

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:31S		
Control 1 EC	0.098%	19:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:32S		
Control 2 EC	0.097%	19:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:33S		
Control 3 EC	0.098%	19:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: \_\_\_\_\_

*LT. F. Pumphrey #3786*

Badge No.: 3786

Date: 01/06/2020