

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00754 Calib. Date: 10/20/2019 Calib. No.: 00029
Certification File No.: 00755 Cert. Date: 10/20/2019 Cert. No.: 00024
Linearity File No.: 00756 Lin. Date: 10/20/2019 Lin. No.: 00024
Solution File No.: 00773 Soln. Date: 01/20/2020 Soln. No.: 00188
Sequential File No.: 00773 File Date: 01/20/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063
Control Solution %: 0.100% Expires: 04/09/2021
Solution Control Lot: 19110 Bottle No.: 0730

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:36S		
Control 1 EC	0.098%	08:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37S		
Control 2 EC	0.098%	08:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38S		
Control 3 EC	0.097%	08:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 01/20/2020