

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00780	Calib. Date: 03/09/2020
Certification File No.:	00781	Calib. No.: 00030
Linearity File No.:	00782	Cert. Date: 03/09/2020
Solution File No.:	00785	Cert. No.: 00025
Sequential File No.:	00785	Lin. Date: 03/09/2020
		Lin. No.: 00025
		Soln. Date: 03/31/2020
		Soln. No.: 00192
		File Date: 03/31/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0064
Solution Control Lot:	19110	Expires: 04/09/2021
		Bottle No.: 0749

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:33S		
Control 1 EC	0.096%	11:33S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:33S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34S		
Control 2 EC	0.097%	11:35S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:35S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:36S		
Control 3 EC	0.097%	11:36S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:36S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 03/31/2020