

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00780	Calib. Date: 03/09/2020
Certification File No.:	00781	Calib. No.: 00030
Linearity File No.:	00782	Cert. Date: 03/09/2020
Solution File No.:	00786	Cert. No.: 00025
Sequential File No.:	00786	Lin. Date: 03/09/2020
		Lin. No.: 00025
		Soln. Date: 04/23/2020
		Soln. No.: 00193
		File Date: 04/23/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0064
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0026

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:16D		
Control 1 EC	0.097%	09:17D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:17D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		
Control 2 EC	0.097%	09:18D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:18D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		
Control 3 EC	0.097%	09:19D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:19D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

L.T. F. Pumphrey #3786

Badge No.: 3786

Date: 04/23/2020