

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE		
Calibration File No.:	00780	Calib. Date:	03/09/2020
Certification File No.:	00781	Cert. No.:	00030
Linearity File No.:	00782	Cert. Date:	03/09/2020
Solution File No.:	00787	Lin. No.:	00025
Sequential File No.:	00787	Soln. Date:	05/21/2020
		Soln. No.:	00194
		File Date:	05/21/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZH-0064
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0090

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:41D		
Control 1 EC	0.097%	18:41D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	18:41D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:42D		
Control 2 EC	0.097%	18:42D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.097%	18:42D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:43D		
Control 3 EC	0.097%	18:44D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.096%	18:44D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: \_\_\_\_\_

*L.T. F. Pumphrey #3786*

Badge No.: 3786

Date: 05/21/2020