

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00870 Calib. Date: 05/13/2022 Calib. No.: 00035
Certification File No.: 00871 Cert. Date: 05/13/2022 Cert. No.: 00030
Linearity File No.: 00872 Lin. Date: 05/13/2022 Lin. No.: 00030
Solution File No.: 00879 Soln. Date: 09/08/2022 Soln. No.: 00229
Sequential File No.: 00879 File Date: 09/08/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063
Control Solution %: 0.100% Expires: 12/28/2023
Solution Control Lot: 21480 Bottle No.: 1469

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:47D		
Control 1 EC	0.101%	14:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:48D		
Control 2 EC	0.100%	14:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50D		
Control 3 EC	0.100%	14:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 09/08/2022