

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE		
Calibration File No.:	00880	Calib. Date:	09/13/2022
Certification File No.:	00881	Cert. Date:	09/13/2022
Linearity File No.:	00882	Lin. Date:	09/13/2022
Solution File No.:	00887	Soln. Date:	12/08/2022
Sequential File No.:	00887	File Date:	12/08/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZH-0064
Solution Control Lot:	22160	Expires:	05/09/2024
		Bottle No.:	1352

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:29S		
Control 1 EC	0.100%	16:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:31S		
Control 2 EC	0.100%	16:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:32S		
Control 3 EC	0.100%	16:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature:

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 12/08/2022