

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXC-0091  
Location: CEDAR GROVE POLICE  
Calibration File No.: 00870 Calib. Date: 05/13/2022 Calib. No.: 00035  
Certification File No.: 00871 Cert. Date: 05/13/2022 Cert. No.: 00030  
Linearity File No.: 00872 Lin. Date: 05/13/2022 Lin. No.: 00030  
Solution File No.: 00875 Soln. Date: 07/04/2022 Soln. No.: 00226  
Sequential File No.: 00875 File Date: 07/04/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063  
Control Solution %: 0.100% Expires: 05/04/2023  
Solution Control Lot: 21150 Bottle No.: 0577

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:23D		
Control 1 EC	0.100%	13:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24D		
Control 2 EC	0.099%	13:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:26D		
Control 3 EC	0.099%	13:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: \_\_\_\_\_

*Capt. F. Pumphrey #3786*

Badge No.: 3786

Date: 07/04/2022