

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXC-0091 |
| Location: | CEDAR GROVE POLICE | |
| Calibration File No.: | 00870 | Calib. Date: 05/13/2022 |
| Certification File No.: | 00871 | Calib. No.: 00035 |
| Linearity File No.: | 00872 | Cert. Date: 05/13/2022 |
| Solution File No.: | 00876 | Cert. No.: 00030 |
| Sequential File No.: | 00876 | Lin. Date: 05/13/2022 |
| | | Lin. No.: 00030 |
| | | Soln. Date: 07/29/2022 |
| | | Soln. No.: 00227 |
| | | File Date: 07/29/2022 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDZH-0063 |
| Solution Control Lot: | 20460 | Expires: 10/21/2022 |
| | | Bottle No.: 1343 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 13:27D | | |
| Control 1 EC | 0.100% | 13:28D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 13:28D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:28D | | |
| Control 2 EC | 0.099% | 13:29D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 13:29D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:30D | | |
| Control 3 EC | 0.099% | 13:30D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 13:30D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:31D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 07/29/2022