

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00870	Calib. Date: 05/13/2022
Certification File No.:	00871	Calib. No.: 00035
Linearity File No.:	00872	Cert. Date: 05/13/2022
Solution File No.:	00878	Cert. No.: 00030
Sequential File No.:	00878	Lin. Date: 05/13/2022
		Lin. No.: 00030
		Soln. Date: 08/12/2022
		Soln. No.: 00228
		File Date: 08/12/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0063
Solution Control Lot:	21080	Expires: 03/08/2023
		Bottle No.: 0511

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:19D		
Control 1 EC	0.099%	13:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20D		
Control 2 EC	0.099%	13:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:22D		
Control 3 EC	0.099%	13:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 08/12/2022