

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00916	Calib. Date: 08/29/2023
Certification File No.:	00917	Calib. No.: 00038
Linearity File No.:	00918	Cert. Date: 08/29/2023
Solution File No.:	00922	Cert. No.: 00033
Sequential File No.:	00922	Lin. Date: 08/29/2023
		Lin. No.: 00033
		Soln. Date: 10/02/2023
		Soln. No.: 00244
		File Date: 10/02/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0063
Solution Control Lot:	23110	Expires: 04/12/2025
		Bottle No.: 1136

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:17D		
Control 1 EC	0.099%	17:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 2 EC	0.099%	17:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		
Control 3 EC	0.098%	17:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 10/02/2023