

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXC-0091  
Location: CEDAR GROVE POLICE  
Calibration File No.: 00916 Calib. Date: 08/29/2023 Calib. No.: 00038  
Certification File No.: 00917 Cert. Date: 08/29/2023 Cert. No.: 00033  
Linearity File No.: 00918 Lin. Date: 08/29/2023 Lin. No.: 00033  
Solution File No.: 00927 Soln. Date: 10/16/2023 Soln. No.: 00245  
Sequential File No.: 00927 File Date: 10/16/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063  
Control Solution %: 0.100% Expires: 04/12/2025  
Solution Control Lot: 23110 Bottle No.: 1279

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:52D		
Control 1 EC	0.095%	08:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 2 EC	0.095%	08:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		
Control 3 EC	0.097%	08:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 10/16/2023