

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00880 Calib. Date: 09/13/2022 Calib. No.: 00036
Certification File No.: 00881 Cert. Date: 09/13/2022 Cert. No.: 00031
Linearity File No.: 00882 Lin. Date: 09/13/2022 Lin. No.: 00031
Solution File No.: 00890 Soln. Date: 02/04/2023 Soln. No.: 00235
Sequential File No.: 00890 File Date: 02/04/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0496

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:43S		
Control 1 EC	0.100%	11:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S		
Control 2 EC	0.099%	11:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46S		
Control 3 EC	0.100%	11:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 02/04/2023