

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00893 Calib. Date: 03/03/2023 Calib. No.: 00037
Certification File No.: 00894 Cert. Date: 03/03/2023 Cert. No.: 00032
Linearity File No.: 00895 Lin. Date: 03/03/2023 Lin. No.: 00032
Solution File No.: 00908 Soln. Date: 05/27/2023 Soln. No.: 00239
Sequential File No.: 00908 File Date: 05/27/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 1353

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:59D		
Control 1 EC	0.099%	20:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00D		
Control 2 EC	0.099%	20:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:02D		
Control 3 EC	0.099%	20:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 05/27/2023