

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE		
Calibration File No.:	00893	Calib. Date:	03/03/2023
Certification File No.:	00894	Cert. No.:	00037
Linearity File No.:	00895	Cert. Date:	03/03/2023
Solution File No.:	00910	Lin. Date:	03/03/2023
Sequential File No.:	00910	Lin. No.:	00032
		Soln. Date:	06/22/2023
		Soln. No.:	00240
		File Date:	06/22/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZH-0063
Solution Control Lot:	22160	Expires:	05/09/2024
		Bottle No.:	0881

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:05D		
Control 1 EC	0.099%	16:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06D		
Control 2 EC	0.098%	16:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08D		
Control 3 EC	0.099%	16:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature:

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 06/22/2023