

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXC-0091  
Location: CEDAR GROVE POLICE  
Calibration File No.: 00893 Calib. Date: 03/03/2023 Calib. No.: 00037  
Certification File No.: 00894 Cert. Date: 03/03/2023 Cert. No.: 00032  
Linearity File No.: 00895 Lin. Date: 03/03/2023 Lin. No.: 00032  
Solution File No.: 00912 Soln. Date: 07/21/2023 Soln. No.: 00241  
Sequential File No.: 00912 File Date: 07/21/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 0547

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:07D		
Control 1 EC	0.100%	10:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		
Control 2 EC	0.100%	10:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 3 EC	0.100%	10:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 07/21/2023