Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE				
Calibration File No.:	00942	Calib. Date:	01/26/2024	Calib. No.:	00039
Certification File No.:	00943	Cert. Date:	01/26/2024	Cert. No.:	00034
Linearity File No.:	00944	Lin. Date:	01/26/2024	Lin. No.:	00034
Solution File No.:	00956	Soln. Date:	05/10/2024	Soln. No.:	00253
Sequential File No.:	00956	File Date:	05/10/2024		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDZH-0064
Control Solution %:	0.100%			Expires:	04/12/2025
Solution Control Lot:	23110			Bottle No.:	1206

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:16D		
Control 1 EC	0.096%	07:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17D		
Control 2 EC	0.097%	07:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	07:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:18D		
Control 3 EC	0.096%	07:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY First Name: FRANCIS MI: M

Signature: Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 05/10/2024