

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXC-0091  
Location: CEDAR GROVE POLICE  
Calibration File No.: 00942 Calib. Date: 01/26/2024 Calib. No.: 00039  
Certification File No.: 00943 Cert. Date: 01/26/2024 Cert. No.: 00034  
Linearity File No.: 00944 Lin. Date: 01/26/2024 Lin. No.: 00034  
Solution File No.: 00957 Soln. Date: 06/10/2024 Soln. No.: 00254  
Sequential File No.: 00957 File Date: 06/10/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064  
Control Solution %: 0.100% Expires: 09/12/2025  
Solution Control Lot: 23330 Bottle No.: 1257

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:59D		
Control 1 EC	0.099%	07:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		
Control 2 EC	0.098%	08:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01D		
Control 3 EC	0.098%	08:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: \_\_\_\_\_

*Capt. F. Pumphrey #3786*

Badge No.: 3786

Date: 06/10/2024