

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00960	Calib. Date: 07/03/2024
Certification File No.:	00961	Calib. No.: 00040
Linearity File No.:	00962	Cert. Date: 07/03/2024
Solution File No.:	00969	Cert. No.: 00035
Sequential File No.:	00969	Lin. Date: 07/03/2024
		Lin. No.: 00035
		Soln. Date: 09/03/2024
		Soln. No.: 00258
		File Date: 09/03/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0063
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 0929

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:23D		
Control 1 EC	0.100%	10:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24D		
Control 2 EC	0.100%	10:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25D		
Control 3 EC	0.100%	10:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 09/03/2024