

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00960	Calib. Date: 07/03/2024
Certification File No.:	00961	Calib. No.: 00040
Linearity File No.:	00962	Cert. Date: 07/03/2024
Solution File No.:	00970	Cert. No.: 00035
Sequential File No.:	00970	Lin. Date: 07/03/2024
		Lin. No.: 00035
		Soln. Date: 10/01/2024
		Soln. No.: 00259
		File Date: 10/01/2024

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDZH-0063
Control Solution %:	0.100%		Expires: 09/12/2025
Solution Control Lot:	23330		Bottle No.: 1103

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:50D		
Control 1 EC	0.099%	08:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 2 EC	0.099%	08:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53D		
Control 3 EC	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 10/01/2024