

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE		
Calibration File No.:	00960	Calib. Date:	07/03/2024
Certification File No.:	00961	Cert. No.:	00040
Linearity File No.:	00962	Cert. Date:	07/03/2024
Solution File No.:	00971	Lin. Date:	07/03/2024
Sequential File No.:	00971	Soln. Date:	10/29/2024
		File Date:	10/29/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZH-0063
Solution Control Lot:	23330	Expires:	09/12/2025
		Bottle No.:	1096

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:47S		
Control 1 EC	0.099%	09:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49S		
Control 2 EC	0.099%	09:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		
Control 3 EC	0.099%	09:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature:

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 10/29/2024