

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXC-0091
Location:	CEDAR GROVE POLICE		
Calibration File No.:	00973	Calib. Date:	11/15/2024
Certification File No.:	00974	Calib. No.:	00041
Linearity File No.:	00975	Cert. Date:	11/15/2024
Solution File No.:	00978	Cert. No.:	00036
Sequential File No.:	00978	Lin. Date:	11/15/2024
		Lin. No.:	00036
		Soln. Date:	12/13/2024
		Soln. No.:	00262
		File Date:	12/13/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZH-0063
Solution Control Lot:	23330	Expires:	09/12/2025
		Bottle No.:	1261

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:48S		
Control 1 EC	0.099%	10:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49S		
Control 2 EC	0.098%	10:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51S		
Control 3 EC	0.099%	10:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 12/13/2024