

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00916 Calib. Date: 08/29/2023 Calib. No.: 00038
Certification File No.: 00917 Cert. Date: 08/29/2023 Cert. No.: 00033
Linearity File No.: 00918 Lin. Date: 08/29/2023 Lin. No.: 00033
Solution File No.: 00940 Soln. Date: 01/10/2024 Soln. No.: 00248
Sequential File No.: 00940 File Date: 01/10/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 1005

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 09:40S | | |
| Control 1 EC | 0.099% | 09:41S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 09:41S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:42S | | |
| Control 2 EC | 0.099% | 09:42S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 09:42S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:43S | | |
| Control 3 EC | 0.099% | 09:44S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 09:44S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:45S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 01/10/2024