

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00973	Calib. Date: 11/15/2024
Certification File No.:	00974	Calib. No.: 00041
Linearity File No.:	00975	Cert. Date: 11/15/2024
Solution File No.:	00983	Cert. No.: 00036
Sequential File No.:	00983	Lin. Date: 11/15/2024
		Lin. No.: 00036
		Soln. Date: 01/10/2025
		Soln. No.: 00263
		File Date: 01/10/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0063
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 0925

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:11S		
Control 1 EC	0.099%	08:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		
Control 2 EC	0.099%	08:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14S		
Control 3 EC	0.099%	08:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 01/10/2025