

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00973	Calib. Date: 11/15/2024
Certification File No.:	00974	Calib. No.: 00041
Linearity File No.:	00975	Cert. Date: 11/15/2024
Solution File No.:	00989	Cert. No.: 00036
Sequential File No.:	00989	Lin. Date: 11/15/2024
		Lin. No.: 00036
		Soln. Date: 01/23/2025
		Soln. No.: 00264
		File Date: 01/23/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0063
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 1258

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:39S		
Control 1 EC	0.098%	08:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40S		
Control 2 EC	0.098%	08:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42S		
Control 3 EC	0.099%	08:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Capt. F. Pumphrey #3786

Badge No.: 3786

Date: 01/23/2025