

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 00973 Calib. Date: 11/15/2024 Calib. No.: 00041
Certification File No.: 00974 Cert. Date: 11/15/2024 Cert. No.: 00036
Linearity File No.: 00975 Lin. Date: 11/15/2024 Lin. No.: 00036
Solution File No.: 01002 Soln. Date: 04/10/2025 Soln. No.: 00267
Sequential File No.: 01002 File Date: 04/10/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0063
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 1212

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:03D		
Control 1 EC	0.097%	08:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04D		
Control 2 EC	0.097%	08:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		
Control 3 EC	0.098%	08:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Chief F. Pumphrey #3786

Badge No.: 3786

Date: 04/10/2025