

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Cedar Grove Police Department 525 Pompton Avenue Cedar Grove, NJ 07009</b>		2. Contact: <b>Lt. Frank Pumphrey</b>	4. Date: <b>01-12-2021</b>
		3. Phone Number: <b>973-239-4100</b>	5. County: <b>Essex</b>
6. Alcotest Instrument Serial Number: <b>n/a</b>	7. Simulator Component Serial Number: <b>DDZH-0063</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Arrived at the department to conduct the bi-annual calibration of Alcotest Instrument.**

10. Comments:  
**Upon the calibration warm up process, it was observed that simulator serial number DDZH-0063 was not turning on. The simulator was taken out of service and will be sent to Draeger for further inspection.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

12. The above Instrument/Component is placed out of service pending further evaluation.

13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Sgr. Marcos Arroyo #6291**  
Name & Badge Number (Print)

*Sgt. M #6291*  
Signature

**01-12-2021**  
Date