

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 01030 Calib. Date: 09/26/2025 Calib. No.: 00043
Certification File No.: 01031 Cert. Date: 09/26/2025 Cert. No.: 00038
Linearity File No.: 01032 Lin. Date: 09/26/2025 Lin. No.: 00038
Solution File No.: 01050 Soln. Date: 11/20/2025 Soln. No.: 00276
Sequential File No.: 01050 File Date: 11/20/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064
Control Solution %: 0.100% Expires: 10/08/2026
Solution Control Lot: 24370 Bottle No.: 0443

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:16S		
Control 1 EC	0.099%	12:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18S		
Control 2 EC	0.099%	12:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19S		
Control 3 EC	0.100%	12:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Francis Pumphrey #3786

Badge No.: 3786

Date: 11/20/2025