

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	01004	Calib. Date: 04/24/2025
Certification File No.:	01005	Calib. No.: 00042
Linearity File No.:	01006	Cert. Date: 04/24/2025
Solution File No.:	01012	Cert. No.: 00037
Sequential File No.:	01012	Lin. Date: 04/24/2025
		Lin. No.: 00037
		Soln. Date: 06/11/2025
		Soln. No.: 00270
		File Date: 06/11/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0064
Solution Control Lot:	24020	Expires: 01/09/2026
		Bottle No.: 0869

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:51D		
Control 1 EC	0.099%	08:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 2 EC	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 3 EC	0.099%	08:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Chloe F. Pumphrey #3786

Badge No.: 3786

Date: 06/11/2025