

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXC-0091
Location: CEDAR GROVE POLICE
Calibration File No.: 01004 Calib. Date: 04/24/2025 Calib. No.: 00042
Certification File No.: 01005 Cert. Date: 04/24/2025 Cert. No.: 00037
Linearity File No.: 01006 Lin. Date: 04/24/2025 Lin. No.: 00037
Solution File No.: 01016 Soln. Date: 07/07/2025 Soln. No.: 00271
Sequential File No.: 01016 File Date: 07/07/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 1199

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:18D		
Control 1 EC	0.098%	15:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 2 EC	0.098%	15:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		
Control 3 EC	0.098%	15:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Chief F. Pumphrey #3786

Badge No.: 3786

Date: 07/07/2025