

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	01030	Calib. Date: 09/26/2025
Certification File No.:	01031	Calib. No.: 00043
Linearity File No.:	01032	Cert. Date: 09/26/2025
Solution File No.:	01039	Cert. No.: 00038
Sequential File No.:	01039	Lin. Date: 09/26/2025
		Lin. No.: 00038
		Soln. Date: 10/24/2025
		Soln. No.: 00275
		File Date: 10/24/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZH-0064
Solution Control Lot:	24370	Expires: 10/08/2026
		Bottle No.: 0446

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:49D		
Control 1 EC	0.101%	13:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		
Control 2 EC	0.100%	13:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51D		
Control 3 EC	0.100%	13:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: Chief F. Pumphrey #3786

Badge No.: 3786

Date: 10/24/2025